

**ANALYZED BY:**

Anresco Laboratories  
1375 Van Dyke Avenue,  
San Francisco, CA 94124  
C8-000052-LIC

**MANUFACTURER:**

3367 25th St  
San Francisco, CA 94110-3924



**SAMPLE INFORMATION**

**Sample No.:** 1102691  
**Product Name:** Toner  
**Matrix:** Topical (Other)  
**Batch #:** GB11122021T

**Date Received:** 11/22/2021  
**Date Reported:** 11/30/2021

**TEST SUMMARY**

<b>Cannabinoid Profile:</b>	✔ Pass	<b>Microbiological Screen:</b>	✔ Pass
<b>Pesticide Residue Screen:</b>	✔ Pass	<b>Residual Solvent Screen:</b>	✔ Pass
<b>Heavy Metal Screen:</b>	✔ Pass	<b>Foreign Material:</b>	✔ Pass
<b>Mycotoxin Screen:</b>	✔ Pass		

**Cannabinoid Profile** ✔ Pass

11/30/2021

**Method:** MF12D012  
**Instrument:** Liquid Chromatography Diode Array Detector (LC-DAD)  
**Limit of Detection** 0.067 mg/g  
**Limit of Quantification** 0.2 mg/g

Cannabinoid	mg/g	%	mg/ml	mg/package	Labeled mg/package	% Difference (mg/package)	Status
Δ8-THC	ND	ND	ND	ND	-	-	-
Δ9-THC	<LOQ	<LOQ	<LOQ	<LOQ	-	-	Pass
Δ9-THCA	ND	ND	ND	ND	-	-	-
THCV	ND	ND	ND	ND	-	-	-
THCVA	ND	ND	ND	ND	-	-	-
CBD	1.10	0.110	1.12	134.31	150	10.46	-
CBDA	ND	ND	ND	ND	-	-	-
CBC	ND	ND	ND	ND	-	-	-
CBCA	ND	ND	ND	ND	-	-	-
CBDV	ND	ND	ND	ND	-	-	-
CBG	ND	ND	ND	ND	-	-	-
CBGA	ND	ND	ND	ND	-	-	-
CBN	ND	ND	ND	ND	-	-	-
Total THC	ND	ND	ND	ND	-	-	-
Total CBD	1.10	0.110	1.12	134.31	-	-	-
Total Cannabinoids	1.10	0.110	1.12	134.31	-	-	-
Total Active Cannabinoids	1.10	0.110	1.12	134.31	-	-	-

**Package Weight (g)** 121.92

**g/ml Conversion Factor** 1.016

**Comments** Confirmed with retest.

**Microbiological Screen** ✔ Pass

11/24/2021

Analyte	Method	Findings	Status
Salmonella	AOAC 2016.01	Negative/25g	Pass
STEC	3M MDS STEC/EAE	Negative/25g	Pass

**Pesticide Residue Screen** ✔ Pass

11/24/2021

**Method:** MF-CHEM-13  
**Instrument:** Liquid Chromatography Tandem Mass Spectrometry (LC-MS/MS) & Gas Chromatography Tandem Mass Spectrometry (GC-MS/MS)

Analyte	LOD/LOQ (µg/g)	Findings (µg/g)	Limit (µg/g)	Status
Abamectin	0.04/0.10	ND	0.3	Pass
Acephate	0.02/0.06	ND	5.0	Pass
Acequinocyl	0.04/0.10	ND	4.0	Pass
Acetamiprid	0.02/0.06	ND	5.0	Pass

Analyte	LOD/LOQ (µg/g)	Findings (µg/g)	Limit (µg/g)	Status
Aldicarb	0.02/0.06	ND	0.02	Pass
Azoxystrobin	0.02/0.06	ND	40.0	Pass
Bifenazate	0.02/0.06	ND	5.0	Pass
Bifenthrin	0.04/0.10	ND	0.5	Pass
Boscalid	0.02/0.06	ND	10.0	Pass
Captan	0.20/0.06	ND	5.0	Pass
Carbaryl	0.02/0.06	ND	0.5	Pass
Carbofuran	0.02/0.06	ND	0.02	Pass
Chlorantraniliprole	0.02/0.06	ND	40.0	Pass
Chlordane	0.02/0.06	ND	0.02	Pass
Chlorfenapyr	0.02/0.08	ND	0.02	Pass
Chlorpyrifos	0.02/0.06	ND	0.02	Pass
Clofentezine	0.02/0.06	ND	0.5	Pass
Coumaphos	0.02/0.06	ND	0.02	Pass
Cyfluthrin	0.10/0.30	ND	1.0	Pass
Cypermethrin	0.10/0.30	ND	1.0	Pass
Daminozide	0.02/0.06	ND	0.02	Pass
DDVP (Dichlorovous)	0.02/0.06	ND	0.02	Pass
Diazinon	0.02/0.06	ND	0.2	Pass
Dimethoate	0.02/0.06	ND	0.02	Pass
Dimethomorph	0.02/0.06	ND	20.0	Pass
Ethoprop(hos)	0.02/0.06	ND	0.02	Pass
Etofenprox	0.02/0.06	ND	0.02	Pass
Etoxazole	0.02/0.06	ND	1.5	Pass
Fenhexamid	0.02/0.06	ND	10.0	Pass
Fenoxycarb	0.02/0.06	ND	0.02	Pass
Fenpyroximate	0.02/0.06	ND	2.0	Pass
Fipronil	0.02/0.06	ND	0.02	Pass
Fonicamid	0.02/0.06	ND	2.0	Pass
Fludioxanil	0.02/0.06	ND	30.0	Pass
Hexythiazox	0.02/0.06	ND	2.0	Pass
Imazalil	0.02/0.06	ND	0.02	Pass
Imidacloprid	0.02/0.06	ND	3.0	Pass
Kresoxim Methyl	0.02/0.06	ND	1.0	Pass
Malathion	0.02/0.06	ND	5.0	Pass
Metalaxyl	0.02/0.06	ND	15.0	Pass
Methiocarb	0.02/0.06	ND	0.02	Pass
Methomyl	0.02/0.06	ND	0.1	Pass
Methyl parathion	0.02/0.06	ND	0.02	Pass
Mevinphos	0.02/0.06	ND	0.02	Pass
Myclobutanil	0.02/0.06	ND	9.0	Pass
Naled	0.02/0.06	ND	0.5	Pass
Oxamyl	0.02/0.06	ND	0.2	Pass
Paclobutrazol	0.02/0.06	ND	0.02	Pass
Pentachloronitrobenzene	0.04/0.10	ND	0.2	Pass
Permethrins	0.10/0.30	ND	20.0	Pass
Phosmet	0.02/0.06	ND	0.2	Pass
Piperonyl Butoxide	0.02/0.06	ND	8.0	Pass
Prallethrin	0.04/0.10	ND	0.4	Pass
Propiconazole	0.02/0.06	ND	20.0	Pass
Propoxur	0.02/0.06	ND	0.02	Pass
Pyrethrins	0.10/0.30	ND	1.0	Pass
Pyridaben	0.02/0.06	ND	3.0	Pass
Spinetoram	0.02/0.06	ND	3.0	Pass
Spinosad	0.02/0.06	ND	3.0	Pass
Spiromesifen	0.04/0.10	ND	12.0	Pass
Spirotetramat	0.02/0.06	ND	13.0	Pass
Spiroxamine	0.02/0.06	ND	0.02	Pass
Tebuconazole	0.02/0.06	ND	2.0	Pass
Thiacloprid	0.02/0.06	ND	0.02	Pass
Thiamethoxam	0.02/0.06	ND	4.5	Pass
Trifloxystrobin	0.02/0.06	ND	30.0	Pass

**Comments** Doubling the LOD/LOQ for Cat II analytes was not necessary as the sample was ND at the listed LOD/LOQ.

**Residual Solvent Screen** ✔ Pass

11/24/2021

**Method:** USP OVI<467>

**Instrument:** Gas Chromatography Mass Spectrometry (GC/MS)

1,2-Dichloroethane	0.10/1.00	ND	1	Pass
--------------------	-----------	----	---	------

Analyte	LOD/LOQ (ppm)	Findings (ppm)	Limit (ppm)	Status
Acetone	50/150	ND	5000	Pass
Acetonitrile	4/12	ND	410	Pass
Benzene	0.10/1.00	ND	1	Pass
n-Butane	48/160	ND	5000	Pass
Chloroform	0.10/1.00	ND	1	Pass
Ethanol	67/200	ND	Not Applicable	Pass
Ethyl acetate	27/80	ND	5000	Pass
Ethyl ether	17/50	ND	5000	Pass
Ethylene oxide	0.50/1.00	ND	1	Pass
n-Heptane	1/4	ND	5000	Pass
n-Hexane	2/10	ND	290	Pass
Isopropyl alcohol	33/100	389.00	Not Applicable	Pass
Methanol	50/150	ND	3000	Pass
Methylene chloride	0.50/1.00	ND	1	Pass
n-Pentane	2/6	ND	5000	Pass
Propane	10/33	ND	5000	Pass
Toluene	10/30	ND	890	Pass
Total xylenes (ortho-, meta-, para-)	30/90	ND	2170	Pass
Trichloroethylene	0.10/1.00	ND	1	Pass

### Heavy Metal Screen ✔ Pass

11/24/2021

Method: MF 24E020

Instrument: Inductively Coupled Plasma Mass Spectrometry (ICP-MS)

Analyte	LOD/LOQ (µg/g)	Findings (µg/g)	Limit (µg/g)	Status
Arsenic	0.02/0.05	ND	1.5	Pass
Cadmium	0.02/0.05	ND	0.5	Pass
Mercury	0.02/0.05	ND	3	Pass
Lead	0.02/0.05	ND	0.5	Pass

### Foreign Material ✔ Pass

11/24/2021

Method: Visual

Analyte	Findings	Limit	Status
Sand, Soils, Cinders, and Dirt	ND	25%	Pass
Mold	ND	25%	Pass
Imbedded Foreign Material	ND	25%	Pass
Insect Fragment	ND	1 per 3g	Pass
Hair	ND	1 per 3g	Pass
Mammalian Excreta	ND	1 per 3g	Pass

### Mycotoxin Screen ✔ Pass

11/24/2021

Method: MF-CHEM-13

Instrument: Liquid Chromatography Tandem Mass Spectrometry (LC-MS/MS) &amp; Gas Chromatography Tandem Mass Spectrometry (GC-MS/MS)

Analyte	LOD/LOQ (µg/kg)	Findings (µg/kg)	Limit (µg/kg)	Status
Aflatoxin B1	2/5	ND	-	-
Aflatoxin B2	2/5	ND	-	-
Aflatoxin G1	2/5	ND	-	-
Aflatoxin G2	2/5	ND	-	-
Total Aflatoxins	8/20	ND	20	Pass
Ochratoxin A	6/20	ND	20	Pass

### Comments

Doubling the LOD/LOQ for Cat II analytes was not necessary as the sample was ND at the listed LOD/LOQ.

(-) = Not Tested, ND = None Detected, &lt;LOQ = Below Limit of Quantitation, LOD = Limit of Detection

Reported by




 Vu Lam  
 Lab Co Director


Scan to verify